

SPECIFICATIONS

| | |
|---------------------|---|
| Model | LD-560 |
| Method | HPLC (High Performance Liquid Chromatography) |
| Parameters | HbA1c, HbF, HbA1a, HbA1b, HbA0; HbA2 (optional) |
| Test Speed | 90 seconds per test |
| Wavelength | 420nm |
| Sample Type | Whole Blood, Pre-diluted blood |
| Autoloader Capacity | 20 tubes |
| Reagents Kit | Eluent A, Eluen B, Hemolysis, Chromatography Column |
| Precision | CV < 2% |
| Column Capacity | 3000 tests |
| Report Range | 3-18% |
| Working Condition | Temperature: 15-30°C Humidity: 30-85% |
| Voltage | 100-240VA |
| Frequency | 50/60Hz |
| Dimensions | 378mm×380mm×510mm |



LABNOVATION TECHNOLOGIES, INC.

+86 755 86368398 +86 755 86368398 www.labnovation.com

Floor 5, Building 1, No. 18 Road, Guangming Hi-Tech Park, Guangming District, Shenzhen 518107, China

LABNOVATION

Most Cost-effective!

LD-560

HPLC System for HbA1c Testing

HPLC System
For HbA1c Testing
With Efficiency



LD-560

HPLC System for HbA1c Testing


Proven golden standard method for HbA1c testing

Fast: 90 seconds per test

CV<2%


Large size touch screen for friendly interface

Most cost-effective




HPLC

- The most recognized principle for HbA1c testing
- CV<2%



Full Automation

- Autoloader for 20 tubes
- Automatically recognizing sample type, and switching the testing mode
- Just one click to start testing



Smart System

- Robust computing system
- Automatic scanning of barcode
- Automatic sample type recognition
- Reagents level alarm



Quality & Performance

- IFCC and NGSP certification
- Control & Calibrators from the original manufacturer



Convenient Operation

- Auto washing for self maintenance at startup
- 10" LCD touch screen to provide rich information
- Easy to change reagents



FACTS ABOUT DIABETES

The number of patients with diabetes increased from 108 million in the year 1980 to 422 million in the year 2014.

In 2014, 8.5% of adults (above age 18) globally has diabetes, and in 1980, this rate was 4.7%. The rate of diabetes is growing faster in the under-developed countries. Diabetes has become the major cause to blindness, kidney failure, heart attack, stroke, and lower limb amputation.

In 2012, 1.5 million of cases of death had direct relation with diabetes; another 2.2 million of death was related with high level of blood sugar. Half of the death with high level of blood sugar happened before their age 70.

According to WHO, diabetes would become the 7th cause of death by 2030.

Glycated Hemoglobin (Hemoglobin A1c, or HbA1c as abbreviation) testing by High Performance Liquid Chromatography (HPLC) method is the golden standard for diabetes diagnosis, and treatment assessment. Every patient with diabetes is advised to have their HbA1c tested for 2-4 times annually. As the education to both the professionals and the patients goes on, the importance of HbA1c testing has been widely accepted and acknowledged.

| LD560 V02.06 | | |
|----------------|------------|-----------------|
| 2019-10-10 | 17:35 | 1-7 |
| ID: 3094159260 | | |
| HbA1c: | 33 | mmol/mol |
| HbA1c: | 5.2 | % |
| HbF: | 1.2 | % |
| | Area | % |
| HbA1a | 0.1 | 1.1 |
| HbA1b | 0.1 | 1.9 |
| HbF | 0.1 | 1.2 |
| LAIc | 0.1 | 1.1 |
| SALc | 0.4 | 5.2 |
| HbA0 | 6.9 | 89.4 |
| Total Area | 7.7 | |

